A Platform to Identify Endogenous Metabolites Using a Novel High Performance Orbitrap MS and the mzCloud Library

Junhua Wang, ¹ David A. Peake, ¹ Robert Mistrik, ² Yingying Huang¹ ¹Thermo Fisher Scientific Inc, San Jose, CA, USA; ²HighChem, Ltd., Bratislava, Slovakia



Overview

Purpose: To demonstrate the capability of the Thermo Scientific™ Orbitrap Fusion™ Tribrid™ Mass Spectrometer in combination with MSn library for metabolite structural identification.

Methods: UHPLC coupled to the Orbitrap Fusion MS to simultaneously perform global profiling and structural elucidation with MSⁿ.

Results: Within UHPLC timescale (2–3 second FWHM), >10 cycles can be finished with fast ion trap HCD MS² for top 11 precursors, each with ion trees of MS²-MS³(top2)-MS⁴(top2).

Introduction

The structural identification of metabolites represents a significant challenge in metabolomics study. Multistage mass spectrometry (MSⁿ) is a powerful tool for compound identification and structural elucidation that goes beyond identifying the exact same compounds, to discovering additional compounds with structural homology to those in the library. Combining with high resolution accurate mass (HRAM) measurement, MSⁿ is highly effective in identifying the unknown but biologically relevant compounds in metabolomics studies[1]. However, the speed on current platforms is yet to be effectively compatible with UHPLC front end, which limited such application in biological samples

Presented here is a new platform to identify endogenous metabolites using a novel high performance Orbitrap hybrid instrument in conjunction with UHPLC and an MS^n library.

Methods

Sample Preparation


Urine samples from an adult male were analyzed. The urine samples were diluted 1 to 10 with water and processed with a $3kD\ MWCO\ filter.$

For plasma samples, proteins were removed with cold methanol and centrifugation, followed by centrifugal evaporation at 35 $^{\circ}$ C. The residue was reconstituted in MeOH/Water, 1/9, and also processed with a 3kD MWCO filter.

Liquid Chromatography


UHPLC separation was implemented on a Dionex Ultimate 3000 HPG (high-pressure gradient) pump using Hypersil GOLD RP C18 column at 450 μ L/min, column temperature at 55 °C. LC solvents were 0.1% FA (A) and 0.1% FA in MeOH (B). A linear gradient was applied from 0.5–50% B for 5.5 min, followed by increasing to 98% at 6 min, held 98% B for 6 min, then decreased to 0.5% at 13min, and equilibrated for another 2 min.

FIGURE 1. Schematic of Orbitrap Fusion Tribrid Mass Spectrometer



An Improved Metabolomics Workflow

A **traditional** workflow for metabolite profiling and identification is based on a full scan LC/MS experiment followed by targeted MS/MS confirmation.



The major problems associated with this methodology are:

- 1) Instrumentation limitations
 - Does not allow simultaneous profiling and MSⁿ in a single run, the scan speed suffers.
 - Identification is usually based on accurate mass and MS², but a large number of LC peaks remain not identifiable.
- 2) MS/MS library limitations
 - MS/MS spectral search is usually manually performed and timeconsuming;
 - · Automated scoring for spectral matching is usually not available;
 - The fragmentation needs to be manually interpreted, which is not only time-consuming but also prone to errors.
- 3) There is no good solution for unknowns.


The Orbitrap Fusion MS enables fast parallel data acquisition on a UHPLC timescale with a variety of dissociation choices

The Orbitrap Fusion MS offers ultimate flexibility in fragmentation methods at all stages of MSⁿ analysis, with fragment ions detected by either mass analyzer. CID-HCD, HCD-CID, and HCD-HCD can all be set up as simple as the conventional CID-CID. This complete flexibility facilitates the examination of fragmentation pathways that leads to the most comprehensive structural information.

mzCloud MSn library enables MSn search for both high and low resolution

mzCloud is a multistage MS library. Based on the spectral tree similarity, an algorithm called precursor ion fingerprinting (PIF) enables the identification of compounds even if they are not in the library, and the fragmental peaks can be annotated [2].

Powered by the novel MS and the mzCloud library, an improved, more connected workflow is shown below:



Major benefits of the improved workflow:

- 1) HCD MS/MS can be performed simultaneously with full scan MS on the UHPLC timescale with ion trap detection at >22 Hz.
- Simultaneous ddMSⁿ ion tree with flexible combinations of CID and/or HCD fragmentation with ion trap and/or Orbitrap™ detection.
- Both the MS² and MSⁿ will be automatically searched in mzCloud, the matching entry will be scored based on PIF, and the fragments will be annotated.
- 4) De novo structural identification is enabled by using this workflow.

Results

FIGURE 1. The Orbitrap Fusion MS and the Drag and Drop Instrument Method Editor.






Metabolite Identification Requires High Resolution and Isotope Ratio Measurements

Structure elucidation of unknown small molecules by mass spectrometry is a great challenge. The first crucial step is to obtain correct elemental compositions.

Natural occurring elements can be monoisotopic (F, Na, P, I) or polyisotopic (H, C, N, O, S, CI, Br) [1]. The molecule L–Methionine $C_5H_{11}NO_2S$ contains multiple polyisotopic elements, especially the S, which result in multiple isotopic peaks needing resolving power above 240,000 at m/z 200 (FWHM). Figure 2A shows the better isotope pattern of molecular ion [M+1+H]*. With increasing resolution up to 450,000 (FWHM), all five isotope peaks are clearly resolved. Figure 2B shows the simulated isotope pattern.


FIGURE 2. Ultrahigh resolution resolves the isotopes



Metabolite Identification and Quantitation in Complex Samples **Requires High Resolution**

D-Glucose C₆H₁₂O₆ and Paraxanthine C₇H₈N₄O₂ co-exist in human blood at distinct concentration levels: 5300 μM (D-Glucose) vs. 10 μM (Paraxanthine) [3]. Although D-Glucose is hundreds of fold more concentrated, it is less readily detectable because of its low ionization efficiency. Their masses differentiate by 0.0013 amu (179.0561 vs. 179.0574), making the detection of these individual components very challenging. Figure 3 shows that, with >300,000 resolving power, the 200 μ M (D-Glucose) and 1 μ M (Paraxanthine) can be readily separated and quantified.

FIGURE 3. Ultrahigh resolution resolves isobaric metabolites



Metabolite Identification Requires High Accuracy of Mass Measurement in Both MS Full Scan Level and MS/MS Level

For an unknown metabolite, the number of calculated chemically possible formulae could be hundreds (20 ppm tolerance window), but the possibilities significantly reduce to tens (3 ppm) and several if the tolerance window is narrowed to sub-ppm.

However, due to the large amount of known isomeric structures, one single formula may correspond to hundreds of structural isomers, thus structure identification of metabolite isomer using only MS full scan is not feasible. MS/ MS and MSn techniques can help solve this analytical problem.

The Orbitrap Fusion MS has a built-in internal calibration functionality (Easy IC). Figure 4 shows LC-MS/MS data of L-Tryptophan from an urine sample with and without internal calibration. Sub-ppm mass accuracy was observed for MS and MS² fragments with IC activated.

FIGURE 4. Easy Internal Calibration





FIGURE 5. Rapid Ion Tree Acquisition on UHPLC Timescale from a Urine Sample (1 MS, 2 MS² and 4 MS³ in 0.6sec/cycle).

(A) Ion trap HCD MS 2 followed by Orbitrap CID MS 3 were acquired for the precursor ion m/z 181.0721. (B) Spectral tree search in mzCloud matches the MS 3 with MS 3 from theophylline with the highest score, indicating the unknown must have the same substructure as theophylline.



Conclusion

The Orbitrap Fusion MS offers ultra high resolution and high fidelity in isotope ratio measurements of metabolites. This allows metabolites in complex mixture with a wide dynamic range be resolved and detected. The Easy-Internal Calibration allows compound assignment be readily done with confidence. The Orbitrap Fusion MS provides ultimate flexibility that facilitates the examination of multiple structural pathways for more structural information at higher speed.

The mzCloud library allows metabolite identification with MS² and/or MSⁿ data in an automated fashion. The identification based on multistage MS spectral trees and PIF algorithm opens the possibility to identify unknown metabolites and isomers. An improved metabolomics workflow is enabled.

References

- 1. Budzikiewicz H, Grigsby RD, Mass Spectrom. Rev. 2006, 25(1):146-157.
- Sheldon MT, Mistrik R, Croley TR, J. Am. Soc. Mass Spectrom. 2009, 20(3): 370–376
- 3. http://www.hmdb.ca/metabolites/HMDB00122 & HMDB01860



©2016 Thermo Fisher Scientific Inc. All rights reserved. All trademarks are the property of Thermo Fisher Scientific, Inc. and its subsidiaries. Specifications, terms and pricing are subject to change. Not all products are available in all countries. Please consult your local sales representative for details.

Africa-Other +27 11 570 1840 Atrica-Other +27 11 570 184 Austrialia +61 3 9757 4300 Austria +43 1 333 50 34 0 Belgium +32 53 73 42 41 Canada +1 800 530 8447 China +86 10 8419 3588 $\textbf{Denmark} \ \ +45\ 70\ 23\ 62\ 60$

Finland/Norway/Sweden +46 8 556 468 00 France +33 1 60 92 48 00 Germany +49 6103 408 1014 India +91 22 6742 9434 Italy +39 02 950 591

Europe-Other +43 1 333 50 34 0 **Japan** +81 45 453 9100 Latin America +1 561 688 8700 Middle East +43 1 333 50 34 0 Netherlands +31 76 579 55 55 New Zealand +64 9 980 6700 Russia/CIS +43 1 333 50 34 0 South Africa +27 11 570 1840

Spain +34 914 845 965 **Switzerland** +41 61 716 77 00 **UK** +44 1442 233555 **USA** +1 800 532 4752



Part of Thermo Fisher Scientific